

Neutron spin echo spectroscopy under 17 T magnetic field at RESEDA

Kindervater, Jonas; Martin, N; Haeussler, Wolfgang; Krautloher, M; Fuchs, C; Muehlbauer, Sebastian; Lim, Joshua; Blackburn, Elizabeth; Boeni, Peter; Pfeleiderer, Christian

DOI:

[10.1051/epjconf/20158303008](https://doi.org/10.1051/epjconf/20158303008)

License:

Creative Commons: Attribution (CC BY)

Document Version

Publisher's PDF, also known as Version of record

Citation for published version (Harvard):

Kindervater, J, Martin, N, Haeussler, W, Krautloher, M, Fuchs, C, Muehlbauer, S, Lim, J, Blackburn, E, Boeni, P & Pfeleiderer, C 2015, 'Neutron spin echo spectroscopy under 17 T magnetic field at RESEDA', *EPJ Web of Conferences*, vol. 83, 03008. <https://doi.org/10.1051/epjconf/20158303008>

[Link to publication on Research at Birmingham portal](#)

Publisher Rights Statement:

©Owned by the authors, published by EDP Sciences, 2015

General rights

Unless a licence is specified above, all rights (including copyright and moral rights) in this document are retained by the authors and/or the copyright holders. The express permission of the copyright holder must be obtained for any use of this material other than for purposes permitted by law.

- Users may freely distribute the URL that is used to identify this publication.
- Users may download and/or print one copy of the publication from the University of Birmingham research portal for the purpose of private study or non-commercial research.
- User may use extracts from the document in line with the concept of 'fair dealing' under the Copyright, Designs and Patents Act 1988 (?)
- Users may not further distribute the material nor use it for the purposes of commercial gain.

Where a licence is displayed above, please note the terms and conditions of the licence govern your use of this document.

When citing, please reference the published version.

Take down policy

While the University of Birmingham exercises care and attention in making items available there are rare occasions when an item has been uploaded in error or has been deemed to be commercially or otherwise sensitive.

If you believe that this is the case for this document, please contact UBIRA@lists.bham.ac.uk providing details and we will remove access to the work immediately and investigate.

Neutron spin echo spectroscopy under 17 T magnetic field at RESEDA

J. Kindervater^{1,a}, N. Martin^{1,2}, W. Häußler^{1,2}, M. Krautloher^{1,2}, C. Fuchs², S. Mühlbauer², J.A. Lim^{2,3,4}, E. Blackburn⁴, P. Böni¹ and C. Pfleiderer¹

¹Physik-Department, Technische Universität München, 85748 Garching, Germany

²Heinz Maier-Leibnitz Zentrum (MLZ), Technische Universität München, 85748 Garching, Germany

³Institut für Festkörperphysik, TU Dresden, 01069 Dresden, Germany

⁴School of Physics and Astronomy, University of Birmingham, Birmingham B15 2TT, UK

Abstract. We report proof-of-principle measurements at the neutron resonance spin echo spectrometer RESEDA (MLZ) under large magnetic fields by means of Modulation of Intensity with Zero Effort (MIEZE). Our study demonstrates the feasibility of applying strong magnetic fields up to 17 T at the sample while maintaining unchanged sub- μeV resolution. We find that the MIEZE-spin-echo resolution curve remains essentially unchanged as a function of magnetic field up to the highest fields available, promising access to high fields without need for additional fine-tuning of the instrument. This sets the stage for the experimental investigations of subtle field dependent phenomena, such as magnetic field-driven phase transitions in hard and soft condensed matter physics.

1. Introduction

A wide range of prominent scientific problems involving high magnetic fields, such as the spectrum of thermal fluctuations stabilising the Skyrmion lattice phase in chiral magnets [1,2], quantum phase transitions of transverse field Ising magnets [3] or field-induced Bose-Einstein condensation of magnons [4], require neutron spectroscopy at sub- μeV resolution. Yet, despite this importance only very few studies of this kind have been reported in the literature. On the one hand, this situation may be traced to the limitations of conventional neutron scattering techniques such as triple axis spectroscopy (TAS) or time of flight spectroscopy (ToF), for which the resolution is directly tied to satisfying strict optical conditions causing a drastic loss of neutron intensity. On the other hand, it has long been established that the necessary high energy resolution may be achieved, in principle, by Neutron Spin Echo (NSE) [5], which, however, requires non-depolarizing samples or sample environments. In turn, it has become an important instrumental challenge to extend neutron spin-echo spectroscopy towards depolarizing samples or sample environments.

As its key idea the NSE technique encodes the information on energy transfers in scattering events by the spin of the neutron. This permits complete decoupling of the energy resolution from the monochromaticity and – to some extent – the divergence of the neutron beam. In turn NSE reaches the highest energy resolution among all neutron spectroscopy techniques reported to date ($\delta E \sim 1 \text{ neV}$), offering a dynamic range of typically 4 to 5 orders of magnitude. On the downside, being based on

polarized neutrons it is crucial that the neutron polarization is not changed in an uncontrolled manner, prohibiting completely depolarizing conditions such as ferromagnetic samples or the application of magnetic fields. In addition strong incoherent scatterers like hydrogenated samples notably reduce the efficiency of classical NSE.

A first approach to overcome the constraints of classical NSE has become known as Ferromagnetic NSE (FNSE) [6] where adiabatic field transitions into the sample regions are used to conserve one polarization component while all others may be lost at the expense of at maximum one half of the polarization. However, due to its inherent complexity, this method, to the best of our knowledge, has only been used very rarely (see *e.g.* [7]). Further, if no component of the neutron beam polarization is conserved and the beam is depolarized completely at the sample, the spin echo signal will be destroyed. A second approach to overcome the limitations of NSE under such completely depolarizing conditions has become known as Intensity Modulation NSE [8]. Representing a straight forward variant of FNSE, two additional spin polarizers before and after the sample are used to transform the phase modulation into an intensity modulation at the sample position. This intensity modulation is unaffected by the depolarization, getting transformed back into a phase modulation behind the sample. However, as its major drawback, this technique is subject to a large additional decrease in efficiency compared to NSE and FNSE. Namely, the two additional polarizers decrease the average intensity I_{av} by at least a factor of two and even by four, if the sample completely depolarizes the beam, and in the ferromagnetic echo the polarization P is, by definition, reduced by another factor of two. Hence, by comparison to NSE this technique is less efficient by at least a

^a e-mail: jonas.kindervater@frm2.tum.de




Figure 1. Principles underlying the MIEZE technique. a) Schematic depiction of a MIEZE-1 setup. b) Kinetic energy of the neutron as function of flight path. c) Temporal separation of the neutron spin wave function as function of the flight path. The MIEZE time τ_{MIEZE} is defined by the separation at the sample position. Figure after [2].

factor of $P^2 \cdot I_{\text{av}} \gtrsim 8 - 16$. In contrast, the technique of Modulation of Intensity with Zero Effort (MIEZE) [9], which is based on the neutron resonance spin echo (NRSE) technique developed by Gähler and Golub [10, 11], allows the combination of spin echo resolution with depolarizing conditions at high intensity. In this paper we report proof-of-principle measurements demonstrating the feasibility of using MIEZE as implemented at the NRSE spectrometer RESEDA (MLZ) for studies under high magnetic fields up to 17 T at the sample position.

2. The MIEZE technique

The MIEZE technique is conceptually based on NRSE, where resonant spin flippers comprised of a static magnetic B_0 and a perpendicular radio frequency field (RF-field) $B_{r,i}$ with frequency ω_i ($i = 1, 2$) induce neutron spin precessions. For a pedagogical introduction we refer to [12]. Shown in Fig. 1a is a schematic depiction of a so-called MIEZE-1 setup consisting of a polarizer, two NRSE spin flippers, a spin analyzer and a time resolving detector. The first NRSE coil (cf. red box in Fig. 1a) introduces a splitting of the kinetic energy of the spin-up (red) and spin-down (blue) state of the neutron as shown in Fig. 1b. The spin states accumulate a difference of time of flight, Δt , due to the difference in kinetic energy, when traveling the distance L_1 in the primary spectrometer arm. The second NRSE coil (cf. green box in Fig. 1a) overcompensates the energy splitting such that this difference of time of flight returns to zero after the distance L_2 at the detector as depicted in Fig. 1c. The MIEZE time τ_{MIEZE} as the difference of time of flight at the sample position, being equivalent to the spin echo time in NSE (cf. Fig. 1a), provides a measure of the resolution of the spectrometer.

Apart from being insensitive to neutron depolarisation the MIEZE technique offers several additional advantages.




Figure 2. Schematic drawing of a TNRSE coil a) and LNRSE coil b). The neutron beam (n) traverses the coils from the left side as indicated by the arrow.

First, being based on two NRSE coils as opposed to four coils in NRSE, the effort required for tuning when setting up the experiment is much reduced in MIEZE. Second, due to its lack of sensitivity on beam depolarisation and when no Bragg peak is accessible in the momentum range of interest the MIEZE technique permits to use strong incoherent scatterers (cf Vanadium or protonated samples) for measurements of the resolution function.

3. Implementation of MIEZE at RESEDA

At the NRSE spectrometer RESEDA (MLZ) the MIEZE option has been implemented in terms of two different variants, a transverse and a longitudinal MIEZE setup referred to as t-MIEZE and l-MIEZE, respectively, which differ in terms of the geometrical arrangement of B_0 and B_r as shown in Fig. 2. In t-MIEZE transverse NRSE (TNRSE) coils are used akin to the standard NRSE setup at RESEDA (cf. Fig. 2a). In our experiments these coils were separated by $L_1 = 2.625$ m. The B_0 -fields in the TNRSE coils are *perpendicular* to the neutron beam and the field region is defined accurately. The region between the TNRSE coils has to be field-free and is therefore shielded by a bespoke mu-metal housing. Coupling coils at the entrance and exit of the mu-metal shield control the transition of the polarization from the guide field of the polarizer into the zero field region and back into the guide field of the analyzer, where a compact V-cavity directly behind the mu-metal shielding is used as analyzer [13]. The t-MIEZE setup covers already now a dynamic range of $0.1 \leq \tau_{\text{MIEZE}} \leq 5$ ns.

In contrast to t-MIEZE the l-MIEZE setup uses a field geometry (LNRSE) [14], where two B_0 -fields are *parallel* to the neutron beam (cf. Fig. 2b). In our experiments the distance was $L_1 = 1.925$ m [15] with the same compact V-cavity as analyzer. The B_0 -field is here generated by two solenoids in a Helmholtz geometry without accurately defined field boundaries. This has the significant advantage that magnetic shielding is not needed and longitudinal guide fields may be used to preserve the polarization. In addition, the similarity of the field geometry to classical spin echo allows to exploit the same correction techniques as in NSE, thus building on well-developed know-how. In its present form the LNRSE already extends the highest spin echo time accessible by a factor of 10 compared to TNRSE. Using effective field integral subtraction [16] the lower limit for the spin echo time is less than 1 ps and the




Figure 3. MIEZE and magnetic field setup at RESEDA. a) Transverse MIEZE setup as combined with an actively shielded 5 T SANS magnet from the MLZ sample environment group. b) Longitudinal MIEZE setup as combined with the Birmingham 17 T magnet [19]. This magnet does not offer active stray-field reduction.

dynamic range in this configuration covers four orders of magnitude.

The secondary spectrometer arm, which is the same for both setups, allows to perform measurements for sample-detector distances between 0.5 and 5 m. A position sensitive CASCADE detector [17, 18] with an active area of 20 cm × 20 cm and a time resolution of $\Delta t = 50$ ns is used. In the CASCADE detector six boron foils (with thicknesses varying from 0.8 to 1.5 μm) convert the neutrons, allowing to record a MIEZE signal with shifted phase on each foil and pixel. Combining the signals originating from all foils offers an efficiency of $\sim 90\%$ of that of a standard ^3He -based detector with a much larger detection depth.

4. Experimental results

For our proof of principle experiment two superconducting magnets have been used. First a cryogen free, 5 T magnet for small angle neutron scattering (SANS) with active stray-field compensation was used. This magnet may be set up for measurements with the magnetic field applied either longitudinal or perpendicular to the neutron beam. Second, a Helium cooled superconducting magnet for longitudinal magnetic fields up to 17 T [19] without stray field compensation. Setting up either magnet at RESEDA, including -slight- retuning the instrument requires approximately half a day.




Figure 4. MIEZE resolution with magnetic field at the sample. a) Typical data as recorded with a 5 T SANS magnet, where the field was applied perpendicular to the neutron beam. b) same as for a) for magnetic field applied longitudinal to the neutron beam. c) Typical data as recorded at 17 T applied longitudinal to the neutron beam using the Birmingham SANS magnet [19].

In a first series of tests the 5 T SANS magnet was used with the t-MIEZE option at a distance of $L_S = 1.38$ m to the detector and a distance between the last TNRSE coil and detector of $L_2 = 3.08$ m. The neutron wavelength was set to $\lambda = 8.33$ Å. The results of a direct beam measurement are shown in Figs. 4a and b. For the application of the field parallel to the neutron beam as combined with the t-MIEZE a solenoid around the spin analyzer was used to compensate the remaining stray fields disturbing the coupling of the neutron spin at the end of the magnetic shielding. In both geometries no reduction of the resolution between zero field and the maximum field of 5 T was observed.

The measurements with the 17 T magnet were performed using the l-MIEZE setup as shown in Fig. 3b. As a general advantage of this setup, the l-MIEZE is less sensitive to the stray field of the magnet as described in further detail below. For our test neutrons with a wavelength of $\lambda = 10.5$ Å were used. The magnet was




Figure 5. Typical MIEZE echo measured with the I-MIEZE setup. The echo was recorded under an applied field of 17 T at a frequency of $\omega_2 = 77$ kHz. The solid line is a cosine fit to the data.

placed at a distance of $L_S = 2.6$ m from the detector and with a distance of $L_2 = 5.00$ m between the last LNRSE coil and the detector. Typical data from direct beam measurements, with 17 T applied to the sample, are shown in Fig. 4c. The scattering of the data points around the single exponential fit is caused by imperfect tuning due to time constraints.

Shown in Fig. 5 is a MIEZE echo at $\tau_{\text{MIEZE}} = 1.12$ ns with 60% contrast as measured at 17 T applied at the location of a sample. In this configuration, the large stray field of the non-compensated longitudinal magnetic field is compensated by a slight modification of the magnitude of the B_0 field produced in the second LNRSE coil. For field integral corrections Pythagoras coils were used. As seen in Fig. 4c, the achieved signal contrast at a Fourier time $\tau_{\text{MIEZE}} \sim 15$ ns is well above the commonly defined resolution limit of $1/e$. An extrapolation towards this limit suggests a resolution for this setup of about 20 ns. In order to reach these high MIEZE times frequencies slightly above 1 MHz and effective magnetic fields of about 70 mT were applied in the LNRSE coils.

5. Conclusions and outlook

In conclusion we have demonstrated that large magnetic fields up to 17 T may readily be combined with the MIEZE technique as implemented at RESEDA at MLZ. As the observed MIEZE resolution is independent from external conditions, field dependent studies are possible without the typical need for fine tuning of the instrument. We expect that these results promise access to a wide range of scientific questions in hard and soft condensed matter by means of high-resolution neutron spectroscopy. Last but not least the possibility to combine high-resolution neutron spectroscopy under high magnetic fields is also of great interest in the form of MIEZE as an add-on option for large scale SANS machines, bridging characteristic times of quasi-elastic measurements and stroboscopic studies in the range $\Delta t \approx 1 \mu\text{s} - 1$ ms in addition to SANS, TISANE and TAS/ToF.

We wish to thank the sample environment group of MLZ, notably J. Peters and H. Weiß, for assistance. Financial support of DFG TRR80 and under ERC advanced grant 291079 are gratefully acknowledged. JK acknowledges financial support by the TUM graduate school. CP acknowledges funding by the BMBF under project 05K10WOC. EB acknowledges support by the EPSRC (EP/G027161/1 and EP/J016977/1). This work was funded by the German BMBF under “Mitwirkung der Zentren der Helmholtz Gemeinschaft und der Technischen Universität München an der Design-Update Phase der ESS, Förderkennzeichen 05E10W01.”

References

- [1] S. Mühlbauer, B. Binz, F. Jonietz, C. Pfleiderer, A. Rosch, A. Neubauer, R. Georgii, P. Böni, *Science* **323**, 915 (2009)
- [2] R. Georgii, B. Brandl, N. Arend, W. Häußler, A. Tischendorf, C. Pfleiderer, P. Böni, J. Lal, *Appl. Phys. Lett.* **98**, 073505 (2011)
- [3] R. Coldea, D.A. Tennant, E.M. Wheeler, E. Wawrzynska, D. Prabhakaran, M. Telling, K. Habicht, P. Smeibidl, K. Kiefer, *Science* **327**, 177 (2010)
- [4] A. Zheludev, V.O. Garlea, T. Masuda, H. Manaka, L.P. Regnault, E. Ressouche, B. Grenier, J.H. Chung, Y. Qiu, K. Habicht et al., *Phys. Rev. B* **76**, 054450 (2007)
- [5] F. Mezei, *Zeitschrift für Physik* **255**, 146 (1972)
- [6] F. Mezei, *The principles of neutron spin echo* (Springer, 1980)
- [7] J. Boucher, F. Mezei, L. Regnault, J. Renard, *Physical review letters* **55**, 1778 (1985)
- [8] B. Farago, F. Mezei, *Physica B + C* **136**, 100 (1986)
- [9] R. Gähler, R. Golub, T. Keller, *Physica B: Condensed Matter* **180**, 899 (1992)
- [10] R. Gähler, R. Golub, *Zeitschrift für Physik B Condensed Matter* **65**, 269 (1987)
- [11] R. Gähler, R. Golub, *Journal de physique* **49**, 1195 (1988)
- [12] T. Keller, R. Golub, R. Gähler, *Neutron spin echo - a technique for high-resolution neutron scattering* (Academic Press, London, 2002), p. 1264
- [13] J. Repper, W. Häußler, A. Ostermann, L. Kredler, A. Chacón, P. Böni, *Phys. Rev. B* **85**, 012036 (2012)
- [14] W. Häußler, U. Schmidt, G. Ehlers, F. Mezei, *Chemical Physics* **292**, 501 (2003)
- [15] M. Krautloher, et al. (in preperation)
- [16] W. Häußler, U. Schmidt, *Physical Chemistry Chemical Physics* **7**, 1245 (2005)
- [17] W. Häußler, P. Böni, M. Klein, C. Schmidt, U. Schmidt, F. Groitl, J. Kindervater, *Review of Scientific Instruments* **82**, 045101 (2011)
- [18] M. Klein, C.J. Schmidt, *Nuclear Instruments and Methods in Physics Research Section A: Accelerators, Spectrometers, Detectors and Associated Equipment* **628**, 9 (2011)
- [19] A.T. Holmes, G.R. Walsh, E. Blackburn, E.M. Forgan, M. Savey-Bennett, *Review of Scientific Instruments* **83**, 023904 (2012)